

Instructions for Use

Life Science Kits & Assays



innuPREP Blood DNA Kit - PP Mini

Order No.:

845-PS-0070016 16 reactions
845-PS-0070096 96 reactions

Publication No.: HB_PS-0070_e_231010

This documentation describes the state at the time of publishing.
It needs not necessarily agree with future versions. Subject to change!

Print-out and further use permitted with indication of source.

© Copyright 2023, IST Innuscreen GmbH

Manufacturer and Distributor:

IST Innuscreen GmbH Phone +49 30 9489 3380
Robert-Rössle-Straße 10 Fax +49 30 9489 3381
13125 Berlin · Germany
Made in Germany! info.innu@ist-ag.com

Contents

1	Introduction.....	2
1.1	Intended use.....	2
1.2	Notes on the use of this manual and the kit.....	3
2	Safety precautions.....	4
3	Storage conditions	5
4	Functional testing and technical assistance.....	5
5	Product use and warranty.....	6
6	Kit components	7
6.1	Components included in the kit.....	7
6.2	Components not included in the kit	7
7	Product specifications	8
8	Initial steps before starting.....	8
9	Sample Preparation	9
9.1	Isolation from 200 µl whole blood samples.....	9
10	Automated extraction using PurePrep Mini	10
10.1	Prefilling of the DW Plate or the DW Strips	10
10.2	Loading filled Deep Well Plate/Strips to the PurePrep Mini and plug in the Tip Combs	10
11	Troubleshooting.....	12

1 Introduction

1.1 Intended use

The innuPREP Blood DNA Kit – PP Mini has been designed for the automated isolation of genomic DNA from whole blood on the PurePrep Mini device. The extraction procedure is based on a new kind of chemistry.

The extraction procedure takes place on the magnetic particle processor PurePrep Mini and allows the parallel and flexible extraction of up to 16 samples.

The extraction process starts with sample lysis on the PurePrep Mini followed by binding of DNA on the surface of the magnetic particle. After washing the DNA is eluted in Elution Buffer.

The kit is intended for use by professional users. The kit has been designed to be used for a wide range of different downstream applications, like amplification reactions and further analytical procedures.



CONSULT INSTRUCTION FOR USE

This package insert must be read carefully prior to use. Package insert instructions must be followed accordingly. Reliability of results cannot be guaranteed if there are any deviations from the instructions in this package insert.

1.2 Notes on the use of this manual and the kit

For easy reference and orientation, the manual and labels use the following warning and information symbols as well as the shown methodology:

Symbol	Information
	REF Catalogue number.
	Content Contains sufficient reagents for <N> reactions.
	Storage conditions Store at room temperature or shown conditions respectively.
	Consult instructions for use This information must be observed to avoid improper use of the kit and the kit components.
	Expiry date
	Lot number The number of the kit charge.
	Manufactured by Contact information of manufacturer.
	For single use only Do not use components for a second time.
	Note / Attention Observe the notes marked in this way to ensure correct function of the kit and to avoid operating errors for obtaining correct results.

The following systematic approach is introduced in the manual:

- The chapters and figures are numbered consecutively.
- A cross reference is indicated with an arrow (e.g. → „Notes on the use of this manual“ p. 3).
- Working steps are numbered.

2 Safety precautions

NOTE

Read through this chapter carefully before use to guarantee your own safety and a trouble-free operation.

Follow all the safety instructions explained in the manual, as well as all messages and information, which are shown.

All due care and attention should be exercised in handling the materials and reagents contained in the kit. Always wear gloves while handling these reagents and avoid any skin contact! In case of contact, flush eyes or skin with a large amount of water immediately.



FOR SINGLE USE ONLY!

This kit is made for single use only!

ATTENTION!

Don't eat or drink components of the kit!

The kit shall only be handled by educated personnel in a laboratory environment!

If the buffer bottles are damaged or leaking, wear gloves and protective goggles when discarding the bottles in order to avoid any injuries. This kit could be used with potential infectious samples. Therefore, all liquid waste must be considered as potentially infectious and must be handled and discarded according to local safety regulation.

Please observe the federal, state and local safety and environmental regulations. Follow the usual precautions for applications using extracted nucleic acids. All materials and reagents used for DNA or RNA isolation should be free of DNases or RNases.

ATTENTION!

Do not add bleach or acidic components to the waste after sample preparation!

NOTE

Emergency medical information in English and German can be obtained 24 hours a day from:

Poison Information Center, Freiburg / Germany

Phone: +49 (0)761 19 240.

For more information on GHS classification and the Safety Data Sheet (SDS) please contact sds.innu@ist-ag.com.

3 Storage conditions

All kit components are shipped at ambient temperature.

Upon arrival, store lyophilized and dissolved **Proteinase K and MAG Suspension F** at 4 °C to 8 °C.

All other components of the kit should be stored dry at room temperature (15 °C to 30 °C). When stored at room temperature, the kit is stable until the expiration date printed on the label on the kit box.

If there are any precipitates within the provided solutions solve these precipitates by careful warming. Before every use make sure that all components have room temperature.

4 Functional testing and technical assistance

The IST Innuscreen GmbH guarantees the correct function of the kit for applications as described in the manual. This product has been produced and tested in an ISO 13485 certified facility.

We reserve the right to change or modify our products to enhance their performance and design. If you have any questions or problems regarding any aspects of the **innuPREP Blood DNA Kit – PP Mini** or other IST Innuscreen GmbH products, please do not hesitate to contact us. For technical support or further information in Germany please contact info.innu@ist.com. For other countries please contact your local distributor.

5 Product use and warranty

The kit is not designed for the usage of other starting materials or other amounts of starting materials than those, referred to in the manual (→ "Product specifications" p. 8). Since the performance characteristics of IST Innuscreen GmbH kits have just been validated for the application described above, the user is responsible for the validation of the performance of IST Innuscreen GmbH kits using other protocols than those described below. IST Innuscreen GmbH kits may be used in clinical diagnostic laboratory systems after the laboratory has validated the complete diagnostic system as required by CLIA' 88 regulations in the U.S. or equivalents in other countries.

All products sold by IST Innuscreen GmbH are subjected to extensive quality control procedures and are warranted to perform as described when used correctly. Any problems should be reported immediately.

NOTE

This kit is for research use only!

6 Kit components

6.1 Components included in the kit

	▽ Σ 16	▽ Σ 96
REF	845-PS-0070016	845-PS-0070096
Lysis Solution CBV (2x)	6 ml	30 ml
MAG Suspension F	0.25 ml	1.1 ml
Binding Solution SBS	8 ml	30 ml
Proteinase K	2 x for 0.3 ml working solution	2 x for 1.5 ml working solution
Washing Solution A	15 ml	90 ml
Washing Solution B2 (conc.)	16 ml	80 ml
Elution Buffer	6 ml	25 ml
Manual	1	1

6.2 Components not included in the kit

- 96 %–99.8 % ethanol (molecular biology grade, undenatured)
- ddH₂O; ultrapure for dissolving Proteinase K
- DW Strip / DW Plate / DW Tip Comb (compatible with the PP Mini device)

7 Product specifications

1. Starting material:
 - Whole blood samples (200 µl)
2. Time for automated extraction protocol on PurePrep Mini:
 - Appr. 58 minutes

8 Initial steps before starting

- Add the indicated volume of ddH₂O to each vial of **Proteinase K**, mix thoroughly and store as described above.

845-PS-0070016	Add 0.3 ml ddH ₂ O to lyophilized Proteinase K.
----------------	--

845-PS-0070096	Add 1.5 ml ddH ₂ O to lyophilized Proteinase K.
----------------	--

- Add the indicated volume of absolute ethanol to **Washing Solution B2 (conc.)** and mix thoroughly. Always keep the bottle firmly closed!

845-PS-0070016	Add 24 ml ethanol to 16 ml Washing Solution B2 .
----------------	---

845-PS-0070096	Add 120 ml ethanol to 80 ml Washing Solution B2 .
----------------	--

9 Sample Preparation

9.1 Isolation from 200 µl whole blood samples

1. Transfer 200 µl of the whole blood sample into the first well of the Deep Well Plate/Strip.
2. Add 300 µl Lysis Solution CBV (2x) and add 20 µl Proteinase K to each well used.

NOTE

If the volume of the blood sample is less than 200 µl adjust with PBS to 200 µl.

3. Proceed with "Automated extraction using PurePrep Mini" on p.10.

10 Automated extraction using PurePrep Mini

10.1 Prefilling of the DW Plate or the DW Strips

Cavity of KF96 DW Plate/Strip	Content
Cavity 1	Sample + 300 µl Lysis Solution CBV (2x) + 20 µl Proteinase K
Cavity 2	800 µl Washing Solution A
Cavity 3	800 µl Washing Solution B2
Cavity 4	800 µl Washing Solution B2
Cavity 5	empty
Cavity 6	100 µl – 200 µl Elution Buffer

10.2 Loading filled Deep Well Plate/Strips to the PurePrep Mini and plug in the Tip Combs

NOTE

- When using strip (strips), the strip is inserted into the tray. In total, a maximum of 8 strips can be used in one extraction-run.
- The modified tip combs always dip staggered into the Strips.
Left tray side: Tip 1, 3, 5, 7
Right tray side: Tip 2, 4, 6, 8.
- It is recommended to mark the tips used for the extraction so that they are not used more than once

-
1. Select the protocol
"BLOODDNA" and start the run.
 2. The automated extraction process starts with sample lysis. After sample lysis the automated run stops.

3. After the device has stopped, take the Plate/Strip out of the device and add 5 μ l of well mixed **MAG Suspension F** and 285 μ l of **Binding Solution SBS** to the lysed samples.

NOTE

Mix the **MAG Solution F** well by vortexing for 1 minute.

4. After addition of **MAG Suspension F** and **Binding Solution SBS** place the Plate/Strip back to the PurePrep Mini and continue the extraction process by starting the device (you will find the instruction on the display of the PurePrep Mini).
5. After finishing the extraction protocol, the Cavity 6 contains the isolated DNA.
6. Transfer the DNA into a fresh 1.5 ml Tube.

IMPORTANT NOTE

After finishing the extraction protocol, the Elution Plate contains the isolated DNA. Store the DNA under adequate conditions.

We recommend storing the extracted DNA for longer use at -22°C to -18°C .

If the eluate contains carryover of magnetic particles, place the plate on a magnet or centrifuge the plate at maximum speed for 3 minutes. Pipet the supernatant with DNA into a new plate.

11 Troubleshooting

Problem / probable cause	Comments and suggestions
Low amount of extracted DNA	
Insufficient lysis	Prolong lysis time. Reduce amount of starting material.
Low concentration of extracted DNA	
Too much Elution Buffer	Elute the DNA in a lower volume of Elution Buffer (min. 80 µl).

IST Innuscreen GmbH
Robert-Rössle-Str.10
13125 Berlin · Germany

Phone +49 30 9489 3380
Fax +49 30 9489 3381

info.innu@ist-ag.com