

Instructions for Use

Life Science Kits & Assays



innuPREP SE Blood direct HMW DNA Kit - KFFLX

Order No.:

845-KF-4196096 96 reactions

845-KF-4196480 480 reactions

Publication No.: HB_KF-4196_e_220818

This documentation describes the state at the time of publishing.
It needs not necessarily agree with future versions. Subject to change!

Print-out and further use permitted with indication of source.

© Copyright 2022, IST Innuscreen GmbH

Manufacturer and distributor
IST Innuscreen GmbH Phone +49 30 9489 3380
Robert-Rössle-Straße 10 Fax +49 30 9489 3381
13125 Berlin · Germany
Made in Germany! info.innu@ist-ag.com

Contents

Contents	1
1 Introduction.....	2
1.1 Intended use.....	2
1.2 Notes on the use of this manual and the kit.....	3
2 Safety precautions.....	4
3 Storage conditions	5
4 Functional testing and technical assistance.....	5
5 Product use and warranty.....	6
6 Kit components	7
6.1 Components included in the kit.....	7
6.2 Components not included in the kit	7
7 Product specifications	8
8 Initial steps before starting.....	8
9 Protocol for isolation of HMW DNA from 200 µl whole blood	9
10 Automated extraction using KingFisher Flex.....	10
10.1 Prefilling of Deep Well Plates	10
10.2 Loading Deep Well Plates to KingFisher Flex	10
10.3 Starting the automated extraction.....	10
11 Troubleshooting.....	12

1 Introduction

1.1 Intended use

The **innuPREP SE Blood direct HMW DNA Kit - KFFLX** has been designed for automated isolation of high molecular weight DNA (HMW) from 200 µl fresh whole blood stabilized with EDTA, citrate or heparin or sampled with PAXgene® Blood DNA Tubes.

The kit is based on the patented SmartExtraction Technology using Smart Modified Surfaces invented by IST Innuscreen GmbH.

The extraction process is based on adsorption of the cells and genomic DNA to Smart Modified Surfaces and it needs no magnetic particles for DNA binding. That means, the DNA binds direct on the surface of the modified KingFisher Flex Tip Combs. After washing, the genomic DNA is eluted from the Smart Modified Surfaces and is ready for use in subsequent downstream applications.

The whole extraction process just needs simple mixing up and down of the modified Tip Combs. The process is very fast and gives no limitation regarding the binding capacity. So, the kit is optimized to get a maximum of yield and quality.











CONSULT INSTRUCTION FOR USE

This package insert must be read carefully before use. Package insert instructions must be followed accordingly. Reliability of results cannot be guaranteed if there are any deviations from the instructions in this package insert.

1.2 Notes on the use of this manual and the kit

For easy reference and orientation, the manual and labels use the following warning and information symbols as well as the shown methodology:

Symbol	Information
	REF Catalogue number.
	Content Contains sufficient reagents for <N> reactions.
	Storage conditions Store at room temperature or shown conditions respectively.
	Consult instructions for use This information must be observed to avoid improper use of the kit and the kit components.
	Expiry date
	Lot number The number of the kit charge.
	Manufactured by Contact information of manufacturer.
	For single use only Do not use components for a second time.
	Note / Attention Observe the notes marked in this way to ensure correct function of the kit and to avoid operating errors for obtaining correct results.

The following systematic approach is introduced in the manual:

- The chapters and figures are numbered consecutively.
- A cross reference is indicated with an arrow (e.g. → „Notes on the use of this manual and the kit“ p. 3).
- Working steps are numbered.

2 Safety precautions

NOTE

Read through this chapter carefully before to guarantee your own safety and a trouble-free operation.

Follow all the safety instructions explained in the manual, as well as all messages and information, which are shown.

All due care and attention should be exercised in handling the materials and reagents contained in the kit. Always wear gloves while handling these reagents and avoid any skin contact! In case of contact, flush eyes or skin with a large amount of water immediately.



FOR SINGLE USE ONLY!

This kit is made for single use only!

ATTENTION!

Don't eat or drink components of the kit!

The kit shall only be handled by educated personnel in a laboratory environment!

If the buffer bottles are damaged or leaking, wear gloves and protective goggles when discarding the bottles in order to avoid any injuries. This kit could be used with potential infectious samples. Therefore, all liquid waste must be considered as potentially infectious and must be handled and discarded according to local safety regulation.

Please observe the federal, state and local safety and environmental regulations. Follow the usual precautions for applications using extracted nucleic acids. All materials and reagents used for DNA or RNA isolation should be free of DNases or RNases.

ATTENTION!

Do not add bleach or acidic components to the waste after sample preparation!

NOTE

Emergency medical information in English and German can be obtained 24 hours a day from:

Poison Information Center, Freiburg / Germany

Phone: +49 (0)761 19 240.

For more information on GHS classification and the Safety Data Sheet (SDS) please contact sds.innu@ist-ag.com.

3 Storage conditions

All kit components are shipped at ambient temperature.

Upon arrival, store lyophilized and dissolved **Proteinase K** at 4 °C to 8 °C and **RNase A** at -22 bis -18°C.

All other components of the **innuPREP SE Blood direct HMW DNA Kit - KFFLX** should be stored dry at room temperature (15 °C to 30 °C). When stored at room temperature, the kit is stable until the expiration date printed on the label on the kit box.

If there are any precipitates within the provided solutions solve these precipitates by careful warming. Before every use make sure that all components have room temperature.

4 Functional testing and technical assistance

The IST Innuscreen GmbH guarantees the correct function of the kit for applications as described in the manual. This product has been produced and tested in an ISO 13485 certified facility.

We reserve the right to change or modify our products to enhance their performance and design. If you have any questions or problems regarding any aspects of the **innuPREP SE Blood direct HMW DNA Kit - KFFLX** or other IST Innuscreen GmbH products, please do not hesitate to contact us. For technical support or further information in Germany please contact info.innu@ist-ag.com. For other countries please contact your local distributor.

5 Product use and warranty

The kit is not designed for the usage of other starting materials or other amounts of starting materials than those, referred to in the manual (→ "Product specifications" p. 8). Since the performance characteristics of IST Innuscreen GmbH kits have just been validated for the application described above, the user is responsible for the validation of the performance of IST Innuscreen GmbH kits using other protocols than those described below. IST Innuscreen GmbH kits may be used in clinical diagnostic laboratory systems after the laboratory has validated the complete diagnostic system as required by CLIA' 88 regulations in the U.S. or equivalents in other countries.

All products sold by IST Innuscreen GmbH are subjected to extensive quality control procedures and are warranted to perform as described when used correctly. Any problems should be reported immediately.

NOTE

The kit is for research use only!

6 Kit components

6.1 Components included in the kit

	Σ 96	Σ 480
REF	845-KF-4196096	845-KF-4196480
Lysis Solution CLS	30 ml	180 ml
Proteinase K	for 2 x 1.5 ml working solution	for 10 x 1.5 ml working solution
RNase A	500 µl	4 x 500 µl
Binding Optimizer	5 ml	2 x 15 ml
Washing Solution S1 (conc.)	20 ml	100 ml
Washing Solution S2 (conc.)	20 ml	100 ml
RNase free Water	25 ml	3 x 30 ml
KF96 Tip Comb with DW Plate	1	5
KF96 modified Tip Comb with DW Plate	1	5
KF96 DW Plate	6	30
Manual	1	1

6.2 Components not included in the kit

- PBS, 1x
- 96 %–99.8 % ethanol (molecular biology grade, undenatured)
- 80 % ethanol for washing plate 3/4
- Isopropanol
- ddH₂O; ultrapure for dissolving Proteinase K

7 Product specifications

1. Starting material:
 - Up to 200µl whole blood (fresh or frozen) treated with EDTA, citrate or heparin
2. Time for automated extraction:
 - 56 minutes

8 Initial steps before starting

- Add the indicated volume of ddH₂O to each vial of Proteinase K, mix thoroughly and store as described above.

845-KF-4096096	Add 1.5 ml ddH ₂ O to lyophilized Proteinase K.
----------------	--

845-KF-4096480	Add 1.5 ml ddH ₂ O to lyophilized Proteinase K.
----------------	--

- Add the indicated amount of ethanol to **Washing Solution S1 (conc.)** and mix thoroughly. Always keep the bottles firmly closed!

845-KF-4096096	Add 80 ml ethanol to 20 ml Washing Solution S1.
----------------	---

845-KF-4096480	Add 400 ml ethanol to 100 ml Washing Solution S1.
----------------	---

- Add the indicated amount of ethanol to **Washing Solution S2 (conc.)** and mix thoroughly. Always keep the bottles firmly closed!

845-KF-4096096	Add 80 ml ethanol to 20 ml Washing Solution S2.
----------------	---

845-KF-4096480	Add 400 ml ethanol to 100 ml Washing Solution S2.
----------------	---

9 Protocol for isolation of HMW DNA from 200 µl whole blood

1. Transfer 200 µl whole blood into a well of the KF96 DW Plate (labeled with "Lysis Plate").
 2. Add 300 µl Lysis Solution CLS and 30 µl Proteinase K
-

NOTE

Optionally add to each sample 5 µl RNase A (10mg/ml).

3. Proceed with "Automated extraction using KingFisher Flex" on p.10.

10 Automated extraction using KingFisher Flex

10.1 Prefilling of Deep Well Plates

Label and fill the Deep Well plates according to the table below.

Plate	Label	Content
Deep Well	Lysis Plate	200µl whole blood, Lysis Solution CLS and Proteinase K described as before
Deep Well	Washing Plate 1	1000 µl Washing Solution S1
Deep Well	Washing Plate 2	1000 µl Washing Solution S2
Deep Well	Washing Plate 3	1000 µl 80% Ethanol
Deep Well	Washing Plate 4	1000 µl 80% Ethanol
Deep Well	Elution Plate	150 µl RNase-free Water
Deep Well	Modified Tip Comb Plate	96 Well Tip Comb <u>modified</u>
Deep Well	Tip Comb Plate	96 Well Tip Comb

10.2 Loading Deep Well Plates to KingFisher Flex

1. Turn on and select the protocol "SE_DNA_1" on KingFisher FLEX instrument and start the run.
2. Follow the instruction and load prefilled Deep Well Plates and Tip Combs successive to the sample tray.

10.3 Starting the automated extraction

1. The automated extraction process starts with sample lysis. After sample lysis the automated run stops.
2. After the device has stopped, take the "Lysis Plate" out of the device.
3. Add 50 µl Binding Optimizer and 400 µl Isopropanol to each sample.

4. After addition of **Binding Optimizer** and **Isopropanol** place the “Lysis Plate” back to the KingFisher Flex and continue the extraction process by starting the KingFisher Flex again (you will find the instruction on the display of the KingFisher Flex).
5. After finishing the extraction protocol, the Elution Plate contains the isolated HMW DNA.

IMPORTANT NOTE HIGH MOLECULAR WEIGHT DNA

The HMW DNA might be very viscous. The dissolving step is crucial for successful extraction and for a maximum of yield. If the DNA content is too high, increase the amount of Elution Buffer.

HMW gDNA needs time to relax. It is generally not recommended to work with freshly eluted DNA unless significant effort is made to ensure even DNA resuspension. Letting a sample relax overnight or for several days facilitates homogenization. If possible, it is recommended that HMW DNA is extracted several days or a week prior to being needed for downstream application.

If you do not need high molecular weight DNA you can shear the DNA e.g. by using ultrasound or by passing the eluate through a needle or a shredder spin filter unit.

11 Troubleshooting

Problem / probable cause	Comments and suggestions
Low amount of extracted DNA	
Preparation without Binding Optimizer and Isopropanol	Pay special attention that Binding Optimizer was added to the lysed sample!
High viscosity of extracted DNA / Inhomogeneous DNA sample	
Relax time too short	Refer to the note of HMW DNA and let the DNA relax overnight at 2-8°C
Degraded or sheared DNA	
Old material insufficient	Old material often contains degraded DNA.
RNA contaminations of extracted DNA	RNase A digestion

IST Innuscreen GmbH
Robert-Rössle-Str.10
13125 Berlin · Germany

Phone +49 30 9489 3380
Fax +49 30 9489 3381

info.innu@ist-ag.com